

šetření. Účastníci si nyní mohou objednávat ze tří skupin kontrolních materiálů (skupina 1 – vzorky E a S po 2 ml, skupina 2 – vzorky E a S po 3 ml a skupina 3 – segmenty K pro zkoušku kompatibility).

Výsledky z roku 2008 a prvních dvou cyklů 2009 potvrzují trendy posledních let – dobrou úspěšnost ABO RhD určení, zvyšující se úspěšnost screeningu (úměrnou zvyšování počtu laboratoří, pracujících citlivějšími a spolehlivějšími technikami sloupcové aglutinace nebo pevné fáze) a ojedinělé problémy při identifikaci protilátek a křížové zkoušce. Titrace protilátek jeví poměrně velký rozptyl výsledků, související s používanými metodami a krvinkami. Pro zbývající cykly byl upraven číselník metod pro titraci a pro zkoušku kompatibility tak, aby bylo hodnocení jednoznačnější. Trvají chyby administrativního charakteru. Do budoucna zvažujeme formální úpravy v zápisu screeningu (u těhotenského screeningu již nebude povinnou součástí enzymový test, proto by současné kategorie byly zavádějící). Kromě kontrolní funkce plní IH cykly externí kontroly i úlohu edukační.

832 washed units in 2004. Total registered side effects could be reduced from 33 to 17 (2004 vs. 2008, $p = < 0,05$). Splitting all side effects to AR, FNHTR and NFNHTR the reduction of the registered events was remarkable for AR (25 vs. 5, 2004 vs. 2008, respectively) and FNHTR (32 vs. 6, 2004 vs. 2008, respectively) whereas the frequency of FNHTR could not be significantly influenced by the use of PAS derived platelets (8 vs. 7, 2004 vs. 2008, respectively). From the analyzed units 351 were transfused mayor incompatible (ABO antibody of the recipient against antigens of the donor), 394 minor incompatible (ABO antibodies in the unit against antigens of the recipient) and 18 mixed incompatible without any registered increase of side effects.

Our data demonstrate a positive effect of the use of PAS on the acceptance of platelet units by the recipients with respect to ABO incompatibility, AR and NFNHTR whereas the number of FNHTR could not be influenced. Especially the acceptance of not ABO matched platelets could simplify the platelet supply for donation services, particularly during weekend and holiday seasons. However, clinical data about the efficacy of PAS derived platelet units, in addition to published data, have to complete our data in the future.

SEKCE V ANGLIČTINĚ

032

PLATELET ADDITIVE SOLUTION- ENHANCED SAFETY AND USABILITY OF TRANSFUSED PLATELET CONCENTRATES

Nussbaumer W.
Innsbruck

Whereas viral infectious risks (HIV, HBV and HCV) of platelet transfusions are nearly negligible today, remaining side effects of platelet transfusions like allergic reactions (AR), febrile (FNHTR) or non febrile non hemolytic transfusion reactions (NFNHTR) as well as bacterial contamination of the platelet concentrate still exists. To overcome these problems the use of platelet additive solutions (PAS) could be of help. Dilution of the produced platelet concentrate with PAS by 66% can reduce the protein content of the unit significantly, dilute therefore the ABO antibody to a level of no harm for the recipient, reduces the risk of allergic side effects and may help to improve the acceptance of the unit by the recipient. Additionally, the usage of PAS is mandatory for nearly all pathogen inactivation methods and reduces the plasma loss of the donor. We decided in 2007 to use PAS for all our produced platelet concentrates and started that policy with 2008. We produce single donor platelets exclusively and 50 % of our platelets are produced on Amicus (Fenwal, Round Zurich, IL, USA) with T-Sol and 50 % were produced with Trima (Caridian, Denver, USA) with SSP+ (Macopharma, Langen, Germany). For both methods, the dilution was 1:3, plasma versus PAS. With consideration of the above mentioned side effects we analyzed all transfused units in 2008 ($n = 3979$) and compared them with the transfused units from 2004 ($n = 3763$), a year all units were produced in 100% plasma. In the past platelets for recipients with repeated side effects during previous transfusions, were washed. Therefore the number of washed units could be a valuable predictor for the acceptance of the platelet units. The number of washed units could be reduced by 86%, which means 117 washed units in 2008 vs.

SEKCE NELÉKAŘSKÝCH PROFESÍ

035

HLEDÁME NOVÉ DÁRCE

Kalužová N., Lakotová O., Čermáková Z.
Krevní centrum Fakultní nemocnice Ostrava

Pro udržení a rozšíření stávající dárcovské základny je jedna z nejdůležitějších věcí „komunikace s klientem“. Dárce by měl z našeho přístupu vycítit, jak je pro nás důležitý a potřebný. Měl by mít zajištěné příjemné prostředí, ve kterém by se cítil dobře. Toto prostředí by měla zajišťovat organizace se vstřícným personálem. Vždyť noví dárči často překonávají velkou bariéru – a tou je strach. Při příchodu klienta na Krevní centrum je vhodná osobní komunikace mezi dárcem a personálem. Na všech úsecích KC by měl zdravotnický personál přívětivě a velice trpělivě vysvětlovat novému dárci jednotlivé kroky, které vedou k darování krve. KC FNO má informace na www.fno.cz. Pravidelná aktualizace internetových stránek a umístění novinek na informačních tabulích KC zajišťuje aktuální přístup k informacím.

Získávání nových dárců krve – „nábor nových dárců“.

Získání nových dárců a udržení stávající dárcovské základny je hlavní náplní práce marketingového pracovníka. Aktivita marketingu KC se orientují zejména na besedy a prezentace na školách a firmách. Marketingový pracovník navštěvuje střední i vysoké školy, kde vysvětluje problematiku darování krve pomocí DVD, které zachycuje cestu dárce od vstupu na KC až po samotné zpracování jeho životadárně tekutiny.

Příkladem těchto aktivit jsou prezentace KC na akcích – „Dny zdraví ve firmách a školách“. Marketingový pracovník může spolupracovat se zdravotními pojišťovnami a ČČK. Součástí strategie jsou také: Propagace v dopravních prostředcích MHD, polep loga KC na autě pro převoz TP, billboard před KC, mediální kampaň v médiích jako je Rádio Kiss Morava, Český rozhlas, Hit Rádio Orion, ČT – pořad „Dobré ráno“.